

| | |
|-------------------------|-----------------------------------|
| Certificate No. | FRT200609 |
| Created on | 09/06/2020 |
| Validity of Certificate | 08/06/ 2022 |
| Part No. | 200605001A – 200605500A |
| Product | Multi Use Temperature Data Logger |
| Quantity | 500 |

FRESHLIANCE ELECTRONICS CORP LTD certifies that the product mentioned hereafter have been produced under ISO9001 : 2008 Quality system, the accuracy and performance is achieved the data described below,

Measuring equipment

high- precision multimeter Thermostatic water cabinet

| Validation points | Validation points | Logger value | Tolerance |
|-------------------|-------------------|--------------|-----------|
| 25°C | 25°C | 25.2°C | ±0.2°C |
| 0°C | 0°C | 0.3°C | ±0.3°C |
| 15°C | 15°C | 15.3°C | ±0.3°C |

Operator

Inspector

Approver

zhao
Jao
Jhu

