



W: www.freshliance.com  
T: +86 371 89908620  
E: sales@freshliance.com

## Manufacturer's Validation Certificate

Certificate No. FRT190820-04  
Created on 20/8/2019  
Validity of Certificate 19/8/ 2020  
Part No. 13071A0032, 13071A0033, 13071A0034, 13071A0035,  
1308130001, 1308130003, 1308130004, 1308130006,  
1308130007, 1308130011, 1308130014, 1308130015,  
1308130021, 1308130022, 1308130026, 1308130028,  
1308130031, 1308130033, 1308130036, 1308130038  
Product Temperature Data Logger  
Quantity 20

FRESHLIANCE ELECTRONICS CORP LTD certifies that the product mentioned hereafter have been produced under ISO9001 : 2008 Quality system, the accuracy and performance is achieved the data described below,

### Measuring equipment

high- precision multimeter


Thermostatic water cabinet

Validation points	Validation points	Logger value	Tolerance
25°C	25°C	25.2°C	±0.2°C
0°C	0°C	0.3°C	±0.3°C
15°C	15°C	15.3°C	±0.3°C

Operator

Inspector

Approver

  
Wang Li Juan  
Li Hong

